



ANATEK LABS

Analytical Results Report For:

City of Mountain Home

Project:

UCMR5

Anatek Work Order:

MFC0669

Anatek Moscow - 1282 Alturas Drive - Moscow, ID 83843 - 208-883-2839 - moscow@anateklabs.com - FL NELAP E87893
Anatek Spokane - 504 E Sprague Ste. D - Spokane, WA 99202 - 509-838-3999 - spokane@anateklabs.com - FL NELAP E871099
Anatek Yakima - 4802 Tielon Drive - Yakima, WA 98908 - 509-225-9404 - yakima@anateklabs.com - FL NELAP E871190
Anatek Wenatchee - 3019 Gs Center Rd - Wenatchee, WA 98801 - 509-701-8362

Anatek Labs, Inc.

1282 Alturas Drive - Moscow, ID 83843 - (208) 883-2839 - email moscow@anateklabs.com
504 E Sprague Ste. D - Spokane, WA 99202 - (509) 838-3999 - email spokane@anateklabs.com

Client: City of Mountain Home
Address: PO Box 10
Mountain Home, ID 83647
Attn: David Sonnentag

Work Order: MFC0669
Project: UCMR5
Reported: 4/29/2025 08:40

Public Drinking Water System UCMR5 Report

PWS Name: **City of Mountain Home** Lab/Sample Number: **MFC0669-01**
Date Received: **03/26/2025** Date Reported by Lab: **04/29/2025**
PWS#: **ID4200032** PWS Facility ID: **13210**
Collect Date: **03/24/2025** Collection Time: **8:45**
Sampling Point ID: **13210TP** Sample Point/ Location: **Well 15**
Sampling Event: **SE1** Sample Type: **FS**
Lab Federal ID#: **ID00013**

Metals Analysis Report:

Analyte	Result	Units	MRL	Analyzed	Analyst	Method
Lithium	ND	ug/L	9.00	4/11/25 13:28	TEC	EPA 200.7

SVOC Analysis Report:

Analyte	Result	Units	MRL	Analyzed	Analyst	Method
11CI-PF3OUds	ND	ug/L	0.0050	4/10/25 19:03	MER	EPA 533
4:2FTS	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
6:2FTS	ND	ug/L	0.0050	4/10/25 19:03	MER	EPA 533
8:2FTS	ND	ug/L	0.0050	4/10/25 19:03	MER	EPA 533
9CI-PF3ONS	ND	ug/L	0.0020	4/10/25 19:03	MER	EPA 533
ADONA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
HFPO-DA	ND	ug/L	0.0050	4/10/25 19:03	MER	EPA 533
NFDHA	ND	ug/L	0.0200	4/10/25 19:03	MER	EPA 533
PFBA	ND	ug/L	0.0050	4/10/25 19:03	MER	EPA 533
PFBS	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFDA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFDaA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFEESA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFHpA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFHpS	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFHxA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFHxS	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFMBA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFMPA	ND	ug/L	0.0040	4/10/25 19:03	MER	EPA 533
PFNA	ND	ug/L	0.0040	4/10/25 19:03	MER	EPA 533
PFOA	ND	ug/L	0.0040	4/10/25 19:03	MER	EPA 533
PFOS	ND	ug/L	0.0040	4/10/25 19:03	MER	EPA 533
PFPeA	ND	ug/L	0.0030	4/10/25 19:03	MER	EPA 533
PFPeS	ND	ug/L	0.0040	4/10/25 19:03	MER	EPA 533
PFUnA	ND	ug/L	0.0020	4/10/25 19:03	MER	EPA 533

Anatek Labs, Inc.

1282 Alturas Drive - Moscow, ID 83843 - (208) 883-2839 - email moscow@anateklabs.com
504 E Sprague Ste. D - Spokane, WA 99202 - (509) 838-3999 - email spokane@anateklabs.com

Client: City of Mountain Home
Address: PO Box 10
Mountain Home, ID 83647
Attn: David Sonnentag

Work Order: MFC0669
Project: UCMR5
Reported: 4/29/2025 08:40

Public Drinking Water System UCMR5 Report

PWS Name: City of Mountain Home Lab/Sample Number: MFC0669-01
Date Received: 03/26/2025 Date Reported by Lab: 04/29/2025
PWS#: ID4200032 PWS Facility ID: 13210
Collect Date: 03/24/2025 Collection Time: 8:45
Sampling Point ID: 13210TP Sample Point/ Location: Well 15
Sampling Event: SE1 Sample Type: FS
Lab Federal ID#: ID00013

SVOC Analysis Report:

Analyte	Recovery	Recovery Limits	Analyzed	Analyst	Method
Surrogate: 13C2-4:2FTS IDA	105%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C2-6:2FTS IDA	98%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C2-8:2FTS IDA	99%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C2-PFDoA IDA	76%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C3-HFPO-DA IDA	78%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C3-PFBS IDA	89%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C3-PFHxS IDA	91%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C4-PFBA IDA	87%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C4-PFHpA IDA	82%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C5-PFHxA IDA	82%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C5-PFPeA IDA	85%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C6-PFDA IDA	83%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C7-PFUnA IDA	80%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C8-PFOA IDA	81%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C8-PFOS IDA	88%	70-130	4/10/25 19:03	MER	EPA 533
Surrogate: 13C9-PFNA IDA	84%	70-130	4/10/25 19:03	MER	EPA 533

Authorized Signature,



Justin Doty For Todd Taruscio, Laboratory Manager

PQL Practical Quantitation Limit
ND Not Detected
MCL EPA's Maximum Contaminant Level
Dry Sample results reported on a dry weight basis
* Not a state-certified analyte

This report shall not be reproduced except in full, without the written approval of the laboratory
The results reported related only to the samples indicated.



ANATEK LABS, INC - Multi-state Certified, NELAC Ac

1282 Alturas Drive, Moscow ID 83843 (208)883-2839 - EPA# ID00
504 E Sprague Ste D, Spokane WA 99202 (509)838-3999 - EPA# W/

UCMR5 Chain of Custody

MFC0669



Due: 04/09/25

WATER SYSTEM	City of Mountain Home	Water System #	42000302
SEND REPORT TO	Devin Schneider	Phone Number	2085872108
ADDRESS	PO Box 101	Email	A.Schneider@cityofmountainhome.com
CITY STATE ZIP	Mountain Home ID 83647	County	ELMORE

PWS FACILITY NAME (e.g., Treatment Plant)	WCI 15	SAMPLING EVENT (circle one)
PWS FACILITY ID (5-digits)	13210 13210	
SAMPLING POINT ID (from inventory)	13210 FB	SE 1 - SE 2 - SE 3 - SE 4

Sampling Point Type Code	FB
Water Source Type Code	<input checked="" type="checkbox"/> GW-Ground Water <input type="checkbox"/> SW-Surface Water <input type="checkbox"/> GU-Ground Water under influence of SW <input type="checkbox"/> MX-Mixed Water

Date and Time Collected	3/24/2025 8:43 AM	Sampler Name	Devin Schneider
		Sampler Signature	[Signature]

Check Desired Analyses

UCMR5 PFAS

EPA 533 EPA 533 Field Reagent Blank transferred from Bottle A to Bottle B

EPA 537.1 EPA 537.1 Field Reagent Blank transferred from Bottle A to Bottle B

UCMR5 Lithium

EPA 200.7 (lithium)

Customer Signature	[Signature]	Received By	FB
Shipping/Delivery Date	3/25/2025	Date/Time Received	3/26/25 9:43



Client Name: c/o Mountain home

TAT: Normal RUSH: _____ days

Samples Received From: FedEx UPS USPS Client Courier Other: _____

Custody Seal on Cooler/Box: Yes No Custody Seals Intact: Yes No N/A

Number of Coolers/Boxes: 1 Type of Ice: Wet Ice Ice Packs Dry Ice None

Packing Material: Bubble Wrap Bags Foam/Peanuts Paper None Other: _____

Cooler Temp As Read (°C): 1.4°C Cooler Temp Corrected (°C): — Thermometer Used: IR-4 IR-6

Samples Received Intact? Yes No N/A
 Chain of Custody Present/Complete? Yes No N/A
 Labels and Chains Agree? Yes No N/A
 Samples Received Within Hold Time? Yes No N/A
 Correct Containers Received? Yes No N/A
 Anatek Bottles Used? Yes No Unknown
 Total Number of Sample Bottles Received:

Comments:

Samples Properly Preserved? Yes No N/A
If No, record preservation and pH-after details
 VOC Vials Free of Headpace (<6mm)? Yes No N/A
 VOC Trip Blanks Present? Yes No N/A

Initial pH:		pH Paper ID:	
<2	or		

Record preservatives (and lot numbers, if known) for containers below:

P250 NH4Ac (2428) 533 x 3
P250 TRIZ (2409) 537 x 3
P250 Li

Notes, comments, etc. (also use this space if contacting the client - record names and date/time)

Received/Inspected By: 703 Date/Time: 3/26/25 9:43
Form F19.01 - Eff 1 Dec 2022